MILLE LACS BAND OF OJIBWE HEALTH AND HUMAN SERVICES POLICY & PROCEDURE

Department: Health Services - Radiological Research	ogy Policy Number: HHS-HS-RAD 1412
Policy Title: Safety Standards for Rad	iology
Attachments:	
Revision History: 07/01, 02/15; 3/18	Revised by/Date: Lance Roeschlein 1/2022
Approved by:	Date:
Jenna Kuduk, Health Services Director	Chunal Belle DN 01/31/22
Approved by:	Date:
Nicole Anderson, Commissioner	ya 2-4-22

POLICY STATEMENT: Minimum standards of safety are utilized in departments using radiographic equipment within the Ne-Ia-Shing, East Lake and Lake Lena Clinics.

PURPOSE: Provide a safe working environment.

PROCEDURE:

- 1. Only authorized persons shall be allowed in rooms with radiation sources.
- 2. Doors to x-ray rooms shall be kept closed whenever machines are energized.
- 3. Warning lights and interlocks shall be installed at all working door entrances to ensure against inadvertent exposure.
- 4. The operator of radiographic equipment shall always be in a shielded position behind the protective wall.
- 5. Critically ill patients shall never be left unattended in a radiographic area.
- 6. If necessary to hold a patient during an exposure, the employee or relative shall be required to adhere to all elements of the Holding policy.
- 7. Provide step-stool to assist patient onto and off the x-ray table.
- 8. No pregnant persons shall assist in restraining a patient during an x-ray exposure.
- 9. The collimator field shall never be larger than clinically necessary.
- 10. A pregnant x-ray technician may not be in the room during an x-ray exposure.
- 11. Signs shall be posted stating "attention, if you are pregnant, inform the technician."
- 12. No persons occupationally exposed to radiation shall routinely hold patients during exposure.
- 13. Appropriate shielding shall be used on all patients to protect them from unnecessary radiation.
- 14. Protective apparel shall be tested for leakage annually.
- 15. Personnel monitoring devices shall be issued to all certified x-ray technicians. In addition, supervisor shall closely monitor all exposure levels on a routine basis.
- 16. Extra equipment shall be kept to a minimum in radiographic room.
- 17. Equipment with mechanical or electrical defects shall be immediately reported to supervisor for corrective action.
- 18. When transporting a patient in a wheelchair, employees shall lock the wheel brake or otherwise secure the vehicle in place before loading or unloading.
- 19. Main switches shall be disconnected at night.

MILLE LACS BAND OF OJIBWE HEALTH AND HUMAN SERVICES POLICY & PROCEDURE

- 20. ALARA rules will be used by radiology staff for all exposures.
- 21. The floors in all areas of the x-ray department will remain dry and clean at all times.
- 22. No food/beverage will be allowed in patient care areas.
- 23. All cassettes will be kept behind the lead in the control area until needed.

Internal and/or External References	Indian Health Manual, Chapter 21 Diagnostic Imaging
	Occupational Safety and Health Administration Standard 1910.132
Compliance - Posting Date	2/3/202244
Replaces - Policy Number	
Next Review - Due Date	